



Axiom™ Genome-Wide BOS 1 Array Plate

Intended Use

The Axiom™ Genome-Wide BOS 1 Array Plate consists of arrays designed for the genotyping of single nucleotide polymorphisms (SNPs). This array plate, when used with the Axiom™ 2.0 Reagent Kit, enables researchers to perform large-scale genotyping studies with minimal hands-on processing per plate.

Each plate consists of 96 microarrays and is offered as a part of a complete automated solution from Affymetrix that includes instrumentation for sample preparation, and array washing, staining and scanning.

The array plate is available in a single configuration:

- 96-Array Plate — Twelve (12) columns of 8 microarrays in each column for a total of 96 identical microarrays on the same plate.

Each microarray on the Axiom Genome-Wide BOS 1 Array Plate enables the genotyping of over 640,000 SNPs. SNPs on this array are derived from the Axiom Bovine Genomic Database including SNPs contributed by the Affymetrix Bovine Community Consortium. The SNPs were discovered by DNA sequencing a variety of *Bos taurus* and *Bos indicus* breeds. SNPs were chosen to maximize genetic coverage in Holstein, Angus, Nelore and related breeds. These SNPs are present on short randomly generated fragments of genomic DNA (25 to 125 base pair), and are amplified using the Axiom 2.0 Reagent Kit. The reagents for processing Axiom arrays are also included in this reagent kit.

SNPs were chosen and represented on the array in part based upon genotyping reproducibility and call rate, with accuracy determined using Mendelian inheritance. These SNPs have been validated across approximately 380 samples covering the five major bovine populations (Taurine beef, Taurine dairy, East Asian Taurine, tropically adapted Taurine and Indicine), which included greater than 200 bovine HapMap samples.

Identical to the cartridge array manufacturing process, the oligonucleotide probes on Axiom Genome-Wide BOS 1 Array Plates are synthesized *in situ* using Affymetrix' photolithographic process.

Reagents, Instrumentation and Software Required

1. Axiom™ 2.0 Reagent Kit
2. Axiom™ Genome-Wide Array Plates
3. Biomek FX^p Target Prep Express by Beckman Coulter
4. GeneTitan® Multi-Channel Instrument
5. Affymetrix GeneChip® Command Console® Software
6. Axiom GW Bos 1 APT Vignettes

For a complete list of reagents and consumables required, please refer to the *Axiom™ 2.0 Assay Automated Workflow Site Prep Guide*, P/N 702984 or the *Axiom™ 2.0 Assay Manual Workflow Site Prep Guide*, P/N 702991.

Library Files

Library files contain information about the probe array design characteristics, probe use and content, and scanning and analysis parameters. These files are unique for each probe array type. Additional information can be located under the specific array product on the Affymetrix web site at www.affymetrix.com/support/technical/libraryfilesmain.affx.

Array Specifications

Axiom Array Specifications	
Number of features	1,416,000
Feature size	5 µm
Length of oligonucleotides	35 nt
Number of probes per SNP	2 or more
Hybridization Volume	105 µL

Instructions for Use

Refer to the following documents for instructions on processing samples using the Axiom™ 2.0 Genotyping Assay.

- *Axiom™ 2.0 Assay Automated Workflow User Guide*, P/N 702963
- *Axiom™ 2.0 Assay Manual Workflow User Guide*, P/N 702990
- *Axiom™ 2.0 Assay Automated Workflow Site Prep Guide*, P/N 702984
- *Axiom™ 2.0 Assay Manual Workflow Site Prep Guide*, P/N 702991
- *Axiom™ 2.0 Assay Automated Target Prep Protocol QRC*, P/N 702962
- *Axiom™ 2.0 Manual Target Prep Protocol QRC*, P/N 702989
- *GeneTitan® MC Protocol for Axiom 2.0 Array Plate Processing QRC*, P/N 702988
- *Axiom™ Genotyping Solution Analysis Guide*, P/N 702961
- *Axiom™ gDNA Sample Prep for Genome-Wide BOS 1 Array Plate QRC*, P/N 702975

Related Documents

The following documents contain information related to the instruments and software required to perform the Axiom Genotyping Assay:

- *GeneTitan® Multi-Channel Instrument User's Manual*, P/N 08-0306
- *GeneTitan® Multi-Channel Instrument Site Preparation Guide*, P/N 08-0305
- *Affymetrix GeneChip® Command Console® Software User Manual*, P/N 702569
- Axiom GW Bos 1 APT Vignettes
- *Biomek® Liquid Handler User's Manual*, Beckman Coulter P/N 987834
- *Biomek® Software User's Manual*, Beckman Coulter P/N 987835

Storage, Handling and Stability

Axiom Genome-Wide BOS 1 Array Plates should be stored at 2 to 8 °C and must not be frozen. Refer to the expiration date on the package label. Do not use arrays or reagents after the expiration date.

When Handling the Axiom Array Plate

Remove the Array Plate from the pouch with gloved hands. The plate is packaged with a blue plastic base. Do not remove the Array Plate from this protective base, or touch the Array Plate directly. Keep the Array Plate in the protective base at all times, including when placed on the GeneTitan® Multi-Channel instrument.

When Handling the 96 Plate Scan Tray

Remove the 96 Plate Scan Tray (Scan Tray) from the pouch with gloved hands. The Scan Tray is packaged with a black protective base and cover. Keep the Scan Tray in the protective base with the cover at all times prior to loading into the GeneTitan Multi-Channel Instrument. Do not touch the bottom of the Scan Tray directly.

The Scan Tray has protruding guide posts that may be sharp and can puncture the pouch if not handled carefully. Take the necessary precautions to avoid injury.

Ordering Information

Product Name	Description	P/N
BOS1 Genotyping Bundle 1	Kit consists of: <ul style="list-style-type: none">One Axiom Genome-Wide BOS 1 Array PlateGeneTitan® Instrument Consumables¹One Axiom™ 2.0 Reagent Kit (96 reactions)	901791

1. The GeneTitan consumables include one Hyb Tray, one Scan Tray, and five Stain Trays with covers. These consumables are required for processing an Axiom array plate on the GeneTitan® Multi-Channel Instrument.

Affymetrix® products can be purchased directly from Affymetrix in the United States, and many European and Asian countries. For all other territories, please view a list of our distribution partners, which can be located at www.affymetrix.com/site/contact/index.affx.

Precautions

1. AXIOM GENOME-WIDE BOS 1 ARRAY PLATES ARE FOR RESEARCH USE ONLY. NOT FOR USE IN DIAGNOSTIC PROCEDURES.
2. Avoid microbial contamination, which may cause erroneous results.
3. **WARNING: All biological specimens and materials with which they come into contact should be handled as if capable of transmitting infection and disposed of with proper precautions in accordance with federal, state, and local regulations. This includes adherence to the OSHA Bloodborne Pathogens Standard (29 CFR 1910.1030) for blood-derived and other samples governed by this act. Never pipet by mouth. Avoid specimen contact with skin and mucous membranes.**
4. CAUTION: Exercise standard precautions when obtaining, handling, and disposing of potentially carcinogenic reagents.
5. Exercise care to avoid cross-contamination of samples during all steps of this procedure, as this may lead to erroneous results.
6. Use powder-free gloves whenever possible to minimize introduction of powder particles into sample or probe array plates.
7. CAUTION: Use care when handling the Scan Tray as it has protruding guiding posts that may be sharp and can stick out of the pouch if not handled carefully.

Limited License

Subject to the Affymetrix terms and conditions that govern your use of Affymetrix products, Affymetrix grants you a non-exclusive, non-transferable, non-sublicensable license to use this Affymetrix product only in accordance with the manual and written instructions provided by Affymetrix. You understand and agree that except as expressly set forth in the Affymetrix terms and conditions, that no right or license to any patent or other intellectual property owned or licensable by Affymetrix is conveyed or implied by this Affymetrix product. In particular, no right or license is conveyed or implied to use this Affymetrix product in combination with a product not provided, licensed or specifically recommended by Affymetrix for such use.

Contact Information

Affymetrix, Inc.

3420 Central Expressway
Santa Clara, CA 95051 USA
E-mail: support@affymetrix.com
Tel: 1-888-362-2447 (1-888-DNA-CHIP)
Fax: 1-408-731-5441

Affymetrix UK Ltd

Voyager, Mercury Park,
Wycombe Lane, Wooburn Green,
High Wycombe HP10 0HH
United Kingdom
E-mail: supporteurope@affymetrix.com
UK and Others Tel: +44 (0) 1628 552550
France Tel: 0800919505
Germany Tel: 01803001334
Fax: +44 (0) 1628 552585

Affymetrix Japan, K. K.

ORIX Hamamatsucho Bldg, 7F
1-24-8 Hamamatsucho, Minato-ku
Tokyo 105-0013, Japan
Tel: +81-3-6430-4020
Fax: +81-3-6430-4021
E-mail for support: supportjapan@affymetrix.com
E-mail for sales: salesjapan@affymetrix.com

Please visit our web site for international distributor contact information
www.affymetrix.com

Patents

Products may be covered by one or more of the following patents: U.S. Patent Nos. 5,445,934; 5,744,305; 6,399,365; 6,733,977; 7,790,389 and other U.S. or foreign patents. Products are manufactured and sold under license from OGT under 5,700,637 and 6,054,270.

Trademarks

Affymetrix®, Axiom™, MyDesign™, Command Console®, DMET™, GeneAtlas™, GeneChip®, GeneChip-compatible™, GeneTitan®, Genotyping Console™, NetAffx®, and Powered by Affymetrix™ are trademarks or registered trademarks of Affymetrix Inc. All other trademarks are the property of their respective owners.

All other trademarks are the property of their respective owners.

Copyright

© 2010-2011 Affymetrix, Inc. All rights reserved.